

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
35 °C	35.267 °C	35.1 °C	-0.167 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	33.2 %HR	2.691 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	50.5 %HR	0.748 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.110 %HR	67.8 %HR	-2.310 %HR	± 3.0 %HR	± 1.9E+00 %HR