

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
52.65 $\mu\text{m}$	52.80 $\mu\text{m}$	0.15 $\mu\text{m}$	$\pm 1.6 \mu\text{m}$	$\pm$	9.9E-02 $\mu\text{m}$
155.23 $\mu\text{m}$	155.0 $\mu\text{m}$	-0.23 $\mu\text{m}$	$\pm 4.7 \mu\text{m}$	$\pm$	2.4E-01 $\mu\text{m}$
257.12 $\mu\text{m}$	257.2 $\mu\text{m}$	0.08 $\mu\text{m}$	$\pm 7.7 \mu\text{m}$	$\pm$	4.0E-01 $\mu\text{m}$
511.72 $\mu\text{m}$	511.6 $\mu\text{m}$	-0.12 $\mu\text{m}$	$\pm 15.4 \mu\text{m}$	$\pm$	7.8E-01 $\mu\text{m}$