

| <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |              |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|--------------|
| 50.00 mmHg               | 50 mmHg          | 0.00 mmHg            | ± 5 mmHg                    | ±                                  | 6.6E-01 mmHg |
| 100.00 mmHg              | 100 mmHg         | 0.00 mmHg            | ± 5 mmHg                    | ±                                  | 6.5E-01 mmHg |
| 149.00 mmHg              | 150 mmHg         | 1.00 mmHg            | ± 5 mmHg                    | ±                                  | 6.6E-01 mmHg |
| 199.00 mmHg              | 200 mmHg         | 1.00 mmHg            | ± 5 mmHg                    | ±                                  | 6.8E-01 mmHg |
| 249.00 mmHg              | 250 mmHg         | 1.00 mmHg            | ± 5 mmHg                    | ±                                  | 7.5E-01 mmHg |
| 299.00 mmHg              | 300 mmHg         | 1.00 mmHg            | ± 5 mmHg                    | ±                                  | 7.5E-01 mmHg |