

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.453 °C	30.3 °C	-0.153 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	32.5 %HR	2.963 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.366 %HR	51.5 %HR	1.134 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.660 %HR	73.0 %HR	-1.660 %HR	± 3.0 %HR	± 1.9E+00 %HR