

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.453 °C	30.3 °C	-0.153 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	32.1 %HR	2.563 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.366 %HR	51.7 %HR	1.334 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.660 %HR	72.8 %HR	-1.860 %HR	± 3.0 %HR	± 1.9E+00 %HR