

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	30.564 °C	29.7 °C	-0.864 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	29.7 °C	-0.864 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	62.0 %HR	1.777 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	70.448 %HR	74.0 %HR	3.552 %HR	± 5.0 %HR	± 1.9E+00 %HR