

Certificado No. (Certificate #): 66955

Equipo (Instrument): Micrómetro de Exterior

Fecha de Calibración (Calibration Date): 2026-03-11

Marca (Brand): YATO

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.2700 mm	1.270 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
2.5400 mm	2.540 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
3.3020 mm	3.302 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
3.8100 mm	3.810 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
4.0640 mm	4.064 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
4.5720 mm	4.572 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
5.0800 mm	5.080 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
7.6200 mm	7.620 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
10.1600 mm	10.160 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
12.7000 mm	12.700 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.9\text{E-}01 \mu\text{m}$
25.4000 mm	25.400 mm	0 $\mu\text{m}$	$\pm 4 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$