

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
23.30 μm	24 μm	0.70 μm	$\pm 11 \mu\text{m}$	$\pm 1.5\text{E}+00 \mu\text{m}$
51.60 μm	53 μm	1.40 μm	$\pm 12 \mu\text{m}$	$\pm 1.6\text{E}+00 \mu\text{m}$
124.60 μm	127 μm	2.40 μm	$\pm 14 \mu\text{m}$	$\pm 2.1\text{E}+00 \mu\text{m}$
250.30 μm	253 μm	2.70 μm	$\pm 18 \mu\text{m}$	$\pm 1.6\text{E}+00 \mu\text{m}$
486.00 μm	489 μm	3.00 μm	$\pm 25 \mu\text{m}$	$\pm 1.8\text{E}+00 \mu\text{m}$
963.00 μm	969 μm	6.00 μm	$\pm 39 \mu\text{m}$	$\pm 2.2\text{E}+00 \mu\text{m}$