

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.095 °C	1.9 °C	-0.195 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.212 °C	8.1 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.6 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.139 %HR	41.4 %HR	2.261 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.711 %HR	62.8 %HR	-1.911 %HR	± 3.0 %HR	± 1.9E+00 %HR