

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
0 °C	0.481 °C	0.6 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.6 °C	0.064 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.1 °C	0.024 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.0 %HR	0.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	50.3 %HR	0.251 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	80.6 %HR	0.439 %HR	± 5.0 %HR	± 1.9E+00 %HR