

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.253 °C	30.4 °C	0.147 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	33 %HR	2.491 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	54 %HR	4.248 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.223 %HR	82 %HR	6.777 %HR	± 5.0 %HR	± 1.9E+00 %HR