

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.253 °C	30.8 °C	0.547 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	35 %HR	4.491 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	55 %HR	5.248 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.223 %HR	82 %HR	6.777 %HR	± 5.0 %HR	± 1.9E+00 %HR