

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	15.5 °C	-0.022 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.5 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.469 °C	32.4 °C	-0.069 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.3 %HR	2.730 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.0 %HR	-0.448 %HR	± 3.0 %HR	± 1.9E+00 %HR