

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.131 °C	15.1 °C	-0.031 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.259 °C	25.2 °C	-0.059 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.3 °C	-0.113 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.936 %HR	30.4 %HR	0.464 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.906 %HR	59.3 %HR	-0.606 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	79.507 %HR	78.2 %HR	-1.307 %HR	± 5.0 %HR	± 1.9E+00 %HR