

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
12.1 RPM	12 RPM	-0.1 RPM	± 1 RPM	± 1.0E-01	RPM
29.9 RPM	30 RPM	0.1 RPM	± 1 RPM	± 1.0E-01	RPM
59.9 RPM	60 RPM	0.1 RPM	± 1 RPM	± 1.0E-01	RPM