

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0011 ml	1 ml	-0.001 ml	± 0.05 ml	±	7.8E-02 µl
4.9906 ml	5 ml	0.009 ml	± 0.13 ml	±	3.0E-01 µl
10.0086 ml	10 ml	-0.009 ml	± 0.12 ml	±	5.9E-01 µl