

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.5050 ml	0.5 ml	-0.0050 ml	± 0.050 ml	± 6.0E-02 µl
2.9949 ml	3 ml	0.0051 ml	± 0.080 ml	± 2.1E-01 µl
4.991293 ml	5 ml	0.0087 ml	± 0.125 ml	± 3.3E-01 µl