

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.413 °C	29.9 °C	-0.513 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	36.0 %HR	5.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	54.0 %HR	3.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 3.0 %HR	± 1.9E+00 %HR