

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.413 °C	30.5 °C	0.087 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	34.0 %HR	3.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	52.0 %HR	1.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 3.0 %HR	± 1.9E+00 %HR