

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.564 °C	30.8 °C	0.236 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	46.0 %HR	15.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	62.0 %HR	11.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	76.0 %HR	5.552 %HR	± 3.0 %HR	± 1.9E+00 %HR