
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
9.96 RPM	10 RPM	0.04 RPM	± 1 RPM	±	1.9E-01 RPM
19.94 RPM	20 RPM	0.06 RPM	± 1 RPM	±	2.1E-01 RPM
59.94 RPM	60 RPM	0.06 RPM	± 1 RPM	±	2.1E-01 RPM
99.96 RPM	100 RPM	0.04 RPM	± 1 RPM	±	1.8E-01 RPM