

**Clase de Exactitud (Accuracy Class):** Clase II (Class II)

---

<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>	<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.100	0.250	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.200	0.260	± 0.6	± 4.3E-01 mm					
0 a 0.4	0.400	0.260	± 0.7	± 4.3E-01 mm					
0 a 0.6	0.600	0.270	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.800	0.270	± 0.9	± 4.3E-01 mm					
0 a 1	1.000	0.280	± 1	± 4.3E-01 mm					
0 a 1.5	1.500	0.310	± 1.2	± 4.7E-01 mm					
0 a 2	2.000	0.340	± 1.4	± 4.7E-01 mm					
0 a 2.5	2.500	0.380	± 1.6	± 4.7E-01 mm					
0 a 3	3.000	0.430	± 1.8	± 4.7E-01 mm					
0 a 3.5	3.500	0.480	± 2	± 4.7E-01 mm					
0 a 4	4.000	0.540	± 2.2	± 4.7E-01 mm					
0 a 4.5	4.500	0.600	± 2.4	± 4.7E-01 mm					
0 a 5	5.000	0.670	± 2.6	± 4.7E-01 mm					

---