

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.000 mm	1.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
1.250 mm	1.250 mm	0 µm	± 4 µm	±	5.8E-01 µm
1.500 mm	1.500 mm	0 µm	± 4 µm	±	5.8E-01 µm
2.000 mm	2.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
3.000 mm	3.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
5.000 mm	5.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
8.000 mm	8.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
10.000 mm	10.000 mm	0 µm	± 4 µm	±	5.9E-01 µm
15.000 mm	15.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
20.000 mm	20.000 mm	0 µm	± 4 µm	±	5.9E-01 µm
25.000 mm	25.000 mm	0 µm	± 4 µm	±	5.9E-01 µm