

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.481 °C	0.1 °C	-0.381 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.7 °C	0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.1 °C	0.024 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	22.5 %HR	2.155 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	48.9 %HR	-1.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	79.3 %HR	-0.861 %HR	± 5.0 %HR	± 1.9E+00 %HR