

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.481 °C	0.2 °C	-0.281 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.4 °C	-0.136 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.0 °C	-0.076 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.0 %HR	0.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	49.3 %HR	-0.749 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	77.5 %HR	-2.661 %HR	± 5.0 %HR	± 1.9E+00 %HR