

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.403 °C	30.2 °C	-0.203 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	29.599 %HR	31.8 %HR	2.201 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.415 %HR	51.9 %HR	2.485 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	69.854 %HR	72.2 %HR	2.346 %HR	± 2.5 %HR	± 1.9E+00 %HR