

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.30 μm	25 μm	1.50 μm	$\pm 11 \mu\text{m}$	\pm	2.1E+00 μm
51.60 μm	52 μm	0.70 μm	$\pm 12 \mu\text{m}$	\pm	2.2E+00 μm
124.60 μm	126 μm	1.40 μm	$\pm 14 \mu\text{m}$	\pm	2.6E+00 μm
250.30 μm	254 μm	3.70 μm	$\pm 18 \mu\text{m}$	\pm	2.2E+00 μm
486.00 μm	491 μm	5.00 μm	$\pm 25 \mu\text{m}$	\pm	2.4E+00 μm