

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
23.30 μm	25.9 μm	2.63 μm	$\pm 11 \mu\text{m}$	$\pm 2.1\text{E}+00 \mu\text{m}$
51.60 μm	54.9 μm	3.30 μm	$\pm 12 \mu\text{m}$	$\pm 2.2\text{E}+00 \mu\text{m}$
124.60 μm	128.0 μm	3.37 μm	$\pm 14 \mu\text{m}$	$\pm 2.6\text{E}+00 \mu\text{m}$
250.30 μm	255.1 μm	4.77 μm	$\pm 18 \mu\text{m}$	$\pm 2.2\text{E}+00 \mu\text{m}$
486.00 μm	493 μm	7.00 μm	$\pm 25 \mu\text{m}$	$\pm 2.4\text{E}+00 \mu\text{m}$
963.00 μm	974 μm	11.00 μm	$\pm 39 \mu\text{m}$	$\pm 2.6\text{E}+00 \mu\text{m}$