

Resultados (Results) Punto de Rocío

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-12 °C	-12.104 °C	-11.7 °C	0.404 °C	± 2.0 °C	± 4.0E-01 °C
-7 °C	-7.060 °C	-7.9 °C	-0.840 °C	± 2.0 °C	± 4.0E-01 °C
0 °C	-0.381 °C	0.4 °C	0.781 °C	± 2.0 °C	± 4.0E-01 °C