

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)	Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)
0 a 0.1	0.100	0.140	± 0.2	± 4.3E-01 mm					
0 a 0.2	0.200	0.160	± 0.2	± 4.3E-01 mm					
0 a 0.4	0.400	0.200	± 0.2	± 4.3E-01 mm					
0 a 0.6	0.600	0.250	± 0.3	± 4.3E-01 mm					
0 a 0.8	0.800	0.190	± 0.3	± 4.3E-01 mm					
0 a 1	1.000	0.210	± 0.4	± 4.3E-01 mm					
0 a 1.5	1.500	0.250	± 0.5	± 4.7E-01 mm					
0 a 2	2.000	0.320	± 0.6	± 4.7E-01 mm					
0 a 2.5	2.500	0.290	± 0.7	± 4.7E-01 mm					
0 a 3	3.000	0.330	± 0.8	± 4.7E-01 mm					
0 a 3.5	3.500	0.200	± 0.9	± 4.7E-01 mm					
0 a 4	4.000	0.480	± 1	± 4.7E-01 mm					
0 a 4.5	4.500	0.680	± 1.1	± 4.7E-01 mm					
0 a 5	5.000	0.750	± 1.2	± 4.7E-01 mm					
0 a 5.5	5.500	0.840	± 1.3	± 3.4E-01 mm					
0 a 6	6.000	0.920	± 1.4	± 3.4E-01 mm					
0 a 6.5	6.499	0.902	± 1.5	± 3.4E-01 mm					
0 a 7	6.999	0.788	± 1.6	± 3.4E-04 mm					
0 a 7.5	7.499	0.891	± 1.7	± 3.4E-04 mm					
0 a 8	7.999	0.752	± 1.8	± 3.4E-04 mm					