

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
35 °C	35.608 °C	35.4 °C	-0.208 °C	± 0.5 °C	± 4.0E-01 °C
38 °C	38.135 °C	37.8 °C	-0.335 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.040 °C	39.7 °C	-0.340 °C	± 0.5 °C	± 4.0E-01 °C
70 °C	70.183 °C	70.0 °C	-0.183 °C	± 0.5 °C	± 1.1E+00 °C
60 %HR	59.497 %HR	63.6 %HR	4.103 %HR	± 5.0 %HR	± 1.9E+00 %HR
85 %HR	85.448 %HR	82.3 %HR	-3.148 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	90.356 %HR	87.7 %HR	-2.656 %HR	± 5.0 %HR	± 1.9E+00 %HR