

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 35 °C | 35.608 °C | 35.4 °C | -0.208 °C | ± 0.5 °C | ± 1.6E-01 °C |
| 38 °C | 38.135 °C | 37.8 °C | -0.335 °C | ± 0.5 °C | ± 1.6E-01 °C |
| 40 °C | 40.040 °C | 39.7 °C | -0.340 °C | ± 0.5 °C | ± 1.6E-01 °C |
| 70 °C | 70.183 °C | 70.0 °C | -0.183 °C | ± 0.5 °C | ± 1.8E-01 °C |
| 60 %HR | 59.497 %HR | 63.6 %HR | 4.103 %HR | ± 5.0 %HR | ± 1.2E+00 %HR |
| 85 %HR | 85.448 %HR | 82.3 %HR | -3.148 %HR | ± 5.0 %HR | ± 1.1E+00 %HR |
| 90 %HR | 90.356 %HR | 87.7 %HR | -2.656 %HR | ± 5.0 %HR | ± 1.1E+00 %HR |