

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Incertidumbre (Uncertainty)
35 °C	35.608 °C	35.6 °C	-0.008 °C	± 1.6E-01 °C
38 °C	38.135 °C	38.0 °C	-0.135 °C	± 1.6E-01 °C
40 °C	40.040 °C	39.9 °C	-0.140 °C	± 1.6E-01 °C
70 °C	70.183 °C	70.1 °C	-0.083 °C	± 1.8E-01 °C
60 %HR	59.497 %HR	61.7 %HR	2.203 %HR	± 1.2E+00 %HR
85 %HR	85.448 %HR	82.6 %HR	-2.848 %HR	± 1.1E+00 %HR
90 %HR	90.356 %HR	88.5 %HR	-1.856 %HR	± 1.1E+00 %HR