

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Incertidumbre (Uncertainty)</b>
35 °C	35.608 °C	35.5 °C	-0.108 °C	± 1.6E-01 °C
38 °C	38.135 °C	38.1 °C	-0.035 °C	± 1.6E-01 °C
40 °C	40.040 °C	40.0 °C	-0.040 °C	± 1.6E-01 °C
70 °C	70.183 °C	69.9 °C	-0.283 °C	± 1.8E-01 °C
60 %HR	59.497 %HR	62.9 %HR	3.403 %HR	± 1.2E+00 %HR
85 %HR	85.448 %HR	83.8 %HR	-1.648 %HR	± 1.1E+00 %HR
90 %HR	90.356 %HR	89.6 %HR	-0.756 %HR	± 1.1E+00 %HR