

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 130.3 μm                 | 129.8 μm         | -0.5 μm              | ± 6.49 μm                | ±                                  | 5.8E-01 μm |
| 523.0 μm                 | 522.8 μm         | -0.2 μm              | ± 26.14 μm               | ±                                  | 5.8E-01 μm |