

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
14 °C	14.018 °C	14.2 °C	0.182 °C	± 1.0 °C	± 4.0E-01 °C
19 °C	19.032 °C	19.3 °C	0.268 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	56.0 %HR	-4.223 %HR	± 5.0 %HR	± 1.9E+00 %HR