

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.3 °C	0.293 °C	± 1.0 °C	± 4.0E-01 °C
12 °C	12.113 °C	12.5 °C	0.387 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.049 %HR	52.0 %HR	1.951 %HR	± 5.0 %HR	± 1.9E+00 %HR