

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
12.1 μm	14.0 μm	1.9 μm	$\pm 5 \mu\text{m}$	\pm	5.8E-01 μm
24.1 μm	25.0 μm	0.9 μm	$\pm 5 \mu\text{m}$	\pm	5.8E-01 μm
50.6000 μm	51.0000 μm	0.4000 μm	$\pm 5 \mu\text{m}$	\pm	5.8E-01 μm
120.3000 μm	120.6000 μm	0.3000 μm	$\pm 5 \mu\text{m}$	\pm	5.8E-01 μm
246.2000 μm	247.0000 μm	0.8000 μm	$\pm 5 \mu\text{m}$	\pm	5.8E-01 μm