

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.72 μm	0.07 μm	$\pm 1.6 \mu\text{m}$	\pm	5.8E-02 μm
155.23 μm	154.4 μm	-0.83 μm	$\pm 4.7 \mu\text{m}$	\pm	6.2E-02 μm
257.12 μm	257.4 μm	0.28 μm	$\pm 7.7 \mu\text{m}$	\pm	7.0E-02 μm
511.72 μm	512.0 μm	0.28 μm	$\pm 15.4 \mu\text{m}$	\pm	9.7E-02 μm