

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
12.1 $\mu\text{m}$	13.3 $\mu\text{m}$	1.2 $\mu\text{m}$	$\pm 5 \mu\text{m}$	$\pm$	1.3E+00 $\mu\text{m}$
24.1 $\mu\text{m}$	24.7 $\mu\text{m}$	0.6 $\mu\text{m}$	$\pm 5 \mu\text{m}$	$\pm$	8.2E-01 $\mu\text{m}$
50.6000 $\mu\text{m}$	51.3333 $\mu\text{m}$	0.73 $\mu\text{m}$	$\pm 5 \mu\text{m}$	$\pm$	8.2E-01 $\mu\text{m}$
120.3000 $\mu\text{m}$	120.5000 $\mu\text{m}$	0.20 $\mu\text{m}$	$\pm 5 \mu\text{m}$	$\pm$	6.0E-01 $\mu\text{m}$
246.2000 $\mu\text{m}$	247.0000 $\mu\text{m}$	0.80 $\mu\text{m}$	$\pm 5 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$