

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.70 μm	52.98 μm	0.28 μm	$\pm 1.6 \mu\text{m}$	\pm	5.8E-02 μm
155.00 μm	155.0 μm	0.00 μm	$\pm 4.7 \mu\text{m}$	\pm	6.2E-02 μm
257.00 μm	257.2 μm	0.20 μm	$\pm 7.7 \mu\text{m}$	\pm	7.0E-02 μm
512.00 μm	511.6 μm	-0.40 μm	$\pm 15.4 \mu\text{m}$	\pm	9.7E-02 μm