

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-5 °C	-5.335 °C	-5.0 °C	0.335 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.341 °C	0.0 °C	-0.341 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.559 °C	10.0 °C	-0.559 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.0 °C	-0.491 °C	± 1.0 °C	± 4.0E-01 °C