
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
50 RPM	50.0 RPM	0.0 RPM	± 0.1 RPM	±	5.8E-02 RPM
80 RPM	80.0 RPM	0.0 RPM	± 0.1 RPM	±	1.2E-01 RPM
100 RPM	100.0 RPM	0.0 RPM	± 0.1 RPM	±	1.7E-01 RPM
150 RPM	150.0 RPM	0.0 RPM	± 0.1 RPM	±	2.3E-01 RPM
200 RPM	200.0 RPM	0.0 RPM	± 0.1 RPM	±	2.9E-01 RPM