

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.941 °C	15.0 °C	0.059 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.079 °C	25.1 °C	0.021 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.158 °C	32.2 °C	0.042 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.783 %HR	22.9 %HR	3.117 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	51.7 %HR	1.948 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	68.1 %HR	-2.266 %HR	± 5.0 %HR	± 1.9E+00 %HR