

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	14.941 °C	15.3 °C	0.359 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.079 °C	25.4 °C	0.321 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.158 °C	32.5 °C	0.342 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.783 %HR	23.2 %HR	3.417 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	52.1 %HR	2.348 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	67.9 %HR	-2.466 %HR	± 5.0 %HR	± 1.9E+00 %HR