

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.941 °C	15.0 °C	0.059 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.079 °C	25.2 °C	0.121 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.158 °C	32.3 °C	0.142 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.783 %HR	22.6 %HR	2.817 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	51.4 %HR	1.648 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	68.4 %HR	-1.966 %HR	± 5.0 %HR	± 1.9E+00 %HR