

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	14.921 °C	15.1 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.3 °C	-0.150 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.469 °C	32.3 °C	-0.169 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.447 %HR	21.2 %HR	0.753 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	51.8 %HR	1.751 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.0 %HR	-1.243 %HR	± 5.0 %HR	± 1.9E+00 %HR