

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
9.9950 µl	10 µl	0.005 µl	± 0.12 µl	±	5.1E-02 µl
50.0052 µl	50 µl	-0.005 µl	± 0.50 µl	±	5.2E-02 µl
100.060 µl	100 µl	-0.060 µl	± 0.80 µl	±	5.3E-02 µl