

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.9824 µl	#1 µl	0.018 µl	± 0.10 µl	±	5.1E-02 µl
5.0052 µl	#5 µl	-0.005 µl	± 0.10 µl	±	5.1E-02 µl
7.0071 µl	#7 µl	-0.007 µl	± 0.10 µl	±	5.1E-02 µl
9.9091 µl	10 µl	0.091 µl	± 0.10 µl	±	5.2E-02 µl