

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.4 μm	0.2 μm	$\pm 2.5 \mu\text{m}$	\pm	2.1E-01 μm
51.1 μm	51.8 μm	0.7 μm	$\pm 2.5 \mu\text{m}$	\pm	2.1E-01 μm
125.1 μm	123 μm	-2.1 μm	$\pm 3.8 \mu\text{m}$	\pm	2.1E-01 μm
478.0 μm	480 μm	2.0 μm	$\pm 14.3 \mu\text{m}$	\pm	2.2E-01 μm
952 μm	954 μm	2.0 μm	$\pm 28.6 \mu\text{m}$	\pm	2.5E-01 μm